

## Hydrolab Chemicals/Standards

Item Number	Detailed Description	Shelf Life (months)	
		Max	Promised
000534HY	pH Calibration Standard, Powder Pillow Kit, pH 10	12	6
000535HY	pH Calibration Standard, Powder Pillow Kit, pH 7	24	12
000537HY	DO Electrolyte for Clark Cell	24	12
000681HY	pH Calibration Standard, Powder Pillow Kit, pH 4	48	12
002483	Turbidity Calibration Standard, Formazin, 4000 NTU, 500 mL	24	12
002587HY	Ammonium Calibration Standard, Low, 500 mL	12	6
002588HY	Ammonium Calibration Standard, High, 500 mL	12	6
007273	Rhodamine Calibration Standard, 8 oz.	24	12
007306	Turbidity Calibration Standard, StablCal, <0.1 NTU, 500 mL	24	12
007307	Turbidity Calibration Standard, StablCal, 20 NTU, 500 mL	24	12
007308	Turbidity Calibration Standard, StablCal, 100 NTU, 500 mL	24	12
007309	Turbidity Calibration Standard, StablCal, 1000 NTU, 500 mL	24	12
007310	Turbidity Calibration Standard, StablCal, 40 NTU, 1 L	24	12
013610HY	Conductivity Calibration Standard, 0.1 mS/cm, 1 L	12	6
013620HY	Conductivity Calibration Standard, 1.413 mS/cm, 1 L	12	6
013640HY	Conductivity Calibration Standard, 12.856 mS/cm, 1 L	18	12
013650HY	Conductivity Calibration Standard, 47.6 mS/cm, 1 L	18	12
013660HY	pH Calibration Standard, pH 4 Red, 500 mL	48	12
013670HY	pH Calibration Standard, pH 7 Yellow, 500 mL	24	12
013680HY	pH Calibration Standard, pH 10 Blue, 500 mL	12	6
013710HY	Turbidity Calibration Standard, Polymer Bead, 50 NTU, 1 L	12	6
013720HY	Turbidity Calibration Standard, Polymer Bead, 100 NTU, 1 L	12	6
013770HY	Conductivity Calibration Standard, 0.5 mS/cm, 1 L	12	6
013800HY	Nitrate Calibration Standard, Low, 4.62 mg/L, 500 mL	12	6
013810HY	Nitrate Calibration Standard, Low, 46.2 mg/L, 500 mL	12	6
013820HY	Chloride Calibration Standard, Low, 500 mL	12	6
013830HY	Chloride Calibration Standard, High, 46.2 mg/L, 500 mL	12	6
013860HY	ORP Calibration Standard, Zobell, 430 mv, 500 mL	6	3